

Seismic Qualification of Krohne Flowmeter

- Introduction
- Test Specimen Identification
- Methodology
- Testing
- Anomalies
- Conclusions

Introduction

RCM Technologies

RCM
Technologies
Canada
Corporation



Krohne Facilities

Test Specimen Identification

- Krohne Variable Area Flowmeter

Model No
H250/RR/M9/K1

Op Pressure
13.8 psig


Op Temp
106°F (41°C)



Test Specimen Identification

(continued)

- Associated Isolated Output Relay Module



Model No.

KFA6-SR2-Ex1.W

Part No. 103374S

Supply 207-253VAC

/45-60 Hz

Input 8 VDC/ 8 mA


Methodology

Step # Description

- 1 Baseline Testing
- 2 Radiation Aging
- 3 Post Radiation Tests
- 4 Thermal Aging
- 5 Post Thermal Aging Tests
- 6 Sine Sweep Seismic Testing
- 7 Post Seismic Tests
- 8 Helium Leak Test

Function Tests

- Test Setups



RCM
Technologies
Canada
Corporation

Radiation Aging


- Flow Meter after Radiation Aging



RCM
Technologies
Canada
Corporation

Thermal Aging

- Flowmeter inside oven
thermocouple →



RCM
Technologies
Canada
Corporation

Thermal Aging (continued)


- Overview of oven



RCM
Technologies
Canada
Corporation

Seismic Testing

- Test set up



RCM
Technologies
Canada
Corporation

Seismic Testing (continued)

RCM
Technologies
Canada
Corporation


RCM Technologies
Mississauga, Ontario

KROHNE

Project # 2251

FIGURE D52
Sine Sweep Seismic Testing
ZY-Orientation
Flowmeter & Relay Module

19 May 2006




Filename: 2006May19-1123-001.vkd

Frequency (Hz)

Test Engineer/Technician:

Seismic Testing (continued)

- Equipment monitoring



RCM
Technologies
Canada
Corporation

Helium Leak Test

- Test set up



RCM
Technologies
Canada
Corporation

Anomalies

- Darkened glass indication cover after Radiation Aging.
- Performance deterioration after Thermal Aging.
- Deflection of needle indicator during Seismic Testing.

Conclusions

RCM
Technologies
Canada
Corporation

